

1. PPE: X-Ray Protective Frontals: Vari-Weight, Vari-Flex, Wide Belt, Standard, Deluxe, Economy, OR Flex, Surgical, and Economy Stretch Belt
X-Ray Protective Wraps: Diamond Back and Stretch Back Reverse
X-Ray Protective Full Overlap Skirt
X-Ray Protective Bolero Vests (Male and Female),
X-Ray Protective Collars: Standard Thyroid, Bib, and Large
X-Ray Protective Small Parts including but not limited to Caps, Sleeves, Shin Guards, Patient Aprons, Blankets, Demi Aprons, Scoliosis Stoles, Breast Shields, Gonad Shields, Gloves, Mittens, Pediatric Aprons
Patient Wraps, Table Drapes, Fixed and Portable Shields
2. The manufacturer and, where applicable, his authorized representative:
Bar-Ray Products, Inc.
90 East Lakeview Drive
Littlestown, PA 17340
3. This declaration of conformity is issued under the sole responsibility of the manufacturer:
Bar-Ray Products, Inc.
90 East Lakeview Drive
Littlestown, PA 17340
4. The objects of this declaration for PPE are as follows:
X-Ray Protective Frontals: Vari-Weight, Vari-Flex, Wide Belt, Standard, Deluxe, Economy, OR Flex, Surgical, and Economy Stretch Belt
X-Ray Protective Wraps: Diamond Back and Stretch Back Reverse
X-Ray Protective Full Overlap Skirt
X-Ray Protective Bolero Vests (Male and Female),
X-Ray Protective Collars: Standard Thyroid, Bib, and Large
X-Ray Protective Small Parts including but not limited to Caps, Sleeves, Shin Guards, Patient Aprons, Blankets, Demi Aprons, Scoliosis Stoles, Breast Shields, Gonad Shields, Gloves, Mittens, Pediatric Aprons
Patient Wraps, Table Drapes, Fixed and Portable Shields
5. The objects of the declaration described in point 4 are in conformity with the provisions of the PPE Regulation (EU) 2016/425.
6. References are also relevant to IEC International Standard 61331-1:2014 and 61331-3:2014.
7. Where applicable, the notified body BTTG, 0338, performed the EU type examination (Module B) and issued the EU type examination Certificate of Conformity No. 523486/1.
8. Where applicable, the PPE is subject to the conformity to type based on the quality assurance of the production process (Module D) under the surveillance of the notified body BTTG, 0338.
9. Additional information: The following core materials are included:
Cost Cruncher (0.25 / 0.35 / 0.50 mm Pb Lead equivalence)
True Lite (0.25 / 0.35 / 0.50 mm Pb Lead equivalence)
Scatter Sentry (0.25 / 0.35 / 0.50 mm Pb Lead equivalence)

Signed for and on behalf of:

Bar-Ray Products, Inc.
90 East Lakeview Drive
Littlestown, PA 17340

By: Carolyn M. Small, Quality Control

Signature

Carolyn M. Small 9/30/20

Date